



**Dutch  
Metrology  
Institute**

## Welcome to session 2: siloxanes content

Joint Workshop of the consortium of “Metrology for biomethane” and ISO/TC193/SC1/WG25 Biomethane

Tuesday 08 September 2020

# Topics

1. NWIP for total silicon concentration (IMBiH)
2. NWIP for siloxanes (NPL)
3. NWIP for halogenated VOCs (VSL + INERIS + RISE)
4. NWIP for HCl + HF (wet chemistry) (INERIS + ISSI)
5. NWIP for ammonia (NPL + VSL + RICE)
6. NWIP for terpenes (RICE + RISE + NPL)
7. NWIP for compressor oil concentration (RISE + INERIS)
8. NWIP for amines concentration (RICE + VSL)
9. NWIP for biogenic methane fraction (RUG)

## Workshop objectives



- Presentation of the methods developed in “Metrology for Biomethane” (EMPIR, 16ENG05) in support of the biomethane specification EN 16723
- Presentation of the validation performed on the methods
- Q&A on the project outcomes
  - Ask questions using the “chat” function in MS Teams
  - You can post your questions any time during the presentation
  - After the presentation, the questions will be addressed
- Sessions are recorded for developing records from the meeting; they will not be made public
- Presentations will be made available on the workshop website

# Development of a standardised test method for siloxanes in biomethane

Lucy Culleton  
Senior Research Scientist, NPL


*16ENG05 Final Workshop, 8<sup>th</sup> September 2020, virtual meeting*

# Need for a NWIP





# Need for a NWIP



# Need for a NWIP



**EN 16723-1**  
**EN 16723-2**



Parameter	EN 16723-1	EN 16723-2
Silicon concentration	$\leq 0.3$ to $1 \text{ mg/m}^3$	$\leq 0.5 \text{ mg/m}^3$
Hydrogen fraction	See EN 16726	$\leq 2 \%$
Hydrocarbon dew point	See EN 16726	$\leq -2 \text{ }^\circ\text{C}$
Oxygen fraction	See EN 16726	$\leq 1 \%$
Sulphur concentration	$\leq 20 \text{ mg/m}^3$	$\leq 5 \text{ mg/m}^3$
Methane number	See EN 16726	$\geq 65$ (80 for high grade)
Compressor oil content	“de minimis”	“de minimis”
Dust impurities	“de minimis”	$\leq 10 \text{ mg/L}$
Amines content	$\leq 10 \text{ mg/m}^3$	$\leq 10 \text{ mg/m}^3$
Water dew point	See EN 16726	$\leq -10 \text{ }^\circ\text{C}$
Chloride concentration	“de minimis”	
Fluoride concentration	“de minimis”	
Carbon monoxide fraction	$\leq 0.1 \%$	
Ammonia concentration	$\leq 10 \text{ mg/m}^3$	


# Siloxanes work within **METROLOGY FOR BIOMETHANE** project

## Aim

To develop **stable**, **metrologically traceable** and **accurate** measurement standards and high-accuracy reference methods for siloxanes in biomethane



Traceable gas standards



Validated measurement methods

## Partners:



## Analysis of natural gas — Biomethane — Determination of siloxane content by Gas Chromatography Ion Mobility Spectrometry

- GC-IMS technique
- Suitable for measurement at EN 16723-1 & EN 16723-2 levels
- Applicable to laboratory and field analysis

# Scope

## Application range:

Component	Abbreviation	Formula	Lower limit (mg m <sup>-3</sup> )	Upper limit (mg m <sup>-3</sup> )
<u>hexamethyldisiloxane</u>	L2	C <sub>6</sub> H <sub>18</sub> Si <sub>2</sub> O	0.1	10.0
<u>octamethyltrisiloxane</u>	L3	C <sub>8</sub> H <sub>24</sub> Si <sub>3</sub> O <sub>2</sub>	0.1	10.0
<u>decamethyltetrasiloxane</u>	L4	C <sub>10</sub> H <sub>30</sub> Si <sub>4</sub> O <sub>3</sub>	0.1	10.0
<u>dodecamethylpentasiloxane</u>	L5	C <sub>12</sub> H <sub>36</sub> Si <sub>5</sub> O <sub>4</sub>	0.1	10.0
<u>hexamethylcyclotrisiloxane</u>	D3	C <sub>12</sub> H <sub>18</sub> O <sub>3</sub> Si <sub>3</sub>	0.1	10.0
<u>octamethylcyclotetrasiloxane</u>	D4	C <sub>8</sub> H <sub>24</sub> O <sub>4</sub> Si <sub>4</sub>	0.1	10.0
<u>octamethylcyclotetrasiloxane</u>	D5	C <sub>10</sub> H <sub>30</sub> O <sub>5</sub> Si <sub>5</sub>	0.1	10.0
<u>dodecamethylcyclohexasiloxane</u>	D6	C <sub>12</sub> H <sub>36</sub> O <sub>6</sub> Si <sub>6</sub>	0.3	10.0

(EN 16723 lower limit specifications are between 0.3 – 1 mg/m<sup>-3</sup>)

# Example instrument for application

A GC-IMS instrument tested within the project

- Linearity
- Trueness
- Repeatability
- Within-lab reproducibility
- Limit of detection
- Selectivity and interferences

Siloxanes tested: L2, L3, D4 and D5



# Note on L3 measurement

- L3 instability noted in certain vessel types

Stability of siloxanes in different vessel types



L2 stability

D4 stability



L3 stability

D5 stability



Agreement between static & dynamic standards




ppm-level

ppb-level

Component	Relative difference in response factor (NPL 3882 vs 40.4 x dilution of NPL 3894)	Expanded uncertainty of measurement ( $k = 2$ )	Relative difference in response factor (NPL 3875 vs 161.8 x dilution of NPL 3894)	Expanded uncertainty of measurement ( $k = 2$ )
L2	0.33 %	1.45 %	5.58 %	1.87 %
L3	0.37 %	1.49 %	-40.58 %	2.11 %
D4	0.10 %	1.66 %	0.96 %	2.14 %
D5	-1.07 %	6.09 %	-3.86 %	4.39 %

# Linearity

- **Dynamic system** used to generate multiple calibration points
- Important to calibrate over **whole range of interest** if response is non-linear



# Trueness

$$\Delta_m = |c_m - c_{ref}|$$

where:

$\Delta_m$  absolute difference between the mean measured value and the reference value from NPL

$c_m$  the mean measured value

$c_{ref}$  the gravimetric standard reference value

- L2, L3 and D4 agreed within 14% of the gravimetric values, which was within the uncertainty of the measurements.
- D5 over-reporting observed. Thought to be due to adsorption effects.
- Dynamic measurement standards may be required for quantifying low level (ppb) L3, due to stability effects in certain cylinder types




# Repeatability & Reproducibility

- Repeatability <3% for all.
- D5 higher value due to adsorption effects

Component:	L2	L3	D4	D5
NPL 3882 s(r) (%)	0.11	0.12	0.20	2.54
NPL 3875 s(r) (%)	0.11	0.38	0.24	0.66

- Method found to be reproduceable <7% for all siloxanes tested

Component:	L2	L3	D4	D5
NPL 3875 RF difference (%)	0.42	-6.46	-1.22	-0.78
NPL 3882 RF difference (%)	0.47	0.10	-0.43	-0.64



# Limit of detection

- Dynamic system used for total Si for measurement.
- Blanks ran using N6.0 methane.
- Limit of detection found to be below the specified EN 16723 levels (0.3 – 1 mg/m<sup>3</sup>).

350 x dilution of NPL 3894	L2	L3	D4	D5	Total silicon (based on D5 value)
LOD (nmol mol <sup>-1</sup> )	12.58	8.33	4.35	1.71	0.63
$U_{\Delta}$ (nmol mol <sup>-1</sup> )	3.22	1.69	1.87	0.74	0.27


Total Si LOD is  
equivalent to  
0.28 mg/m<sup>3</sup>).

# Interferences

- Two mixture types tested

Component	Amount fraction ( $\mu\text{mol mol}^{-1}$ )	
	Mixture 4: NPL A431	Mixture 5: NPL 2059
<i>n</i> -hexane	983.6	-
<i>n</i> -heptane	100.7	-
<i>n</i> -octane	37.8	-
<i>n</i> -nonane	10.5	-
<i>n</i> -decane	5.8	-
Hydrogen sulphide	-	151.4
Ethyl methyl sulphide	-	103.3
Ethyl methyl sulphide	-	102.1
Dimethyl sulphide	-	102.0
Methanethiol	-	100.4
Ethanethiol	-	98.4
Carbonyl sulphide	-	98.2
2-propanethiol	-	98.1
Diethyl sulphide	-	96.1
Balance	Methane	Methane

GC-IMS chromatogram of a C6-C10 hydrocarbon mixture, NPL A431 (red) overlaid with a siloxane mixture NPL3894: 20 x dilution (green).




# Interferences

- Interference noted with sulphur compound

Component	Amount fraction ( $\mu\text{mol mol}^{-1}$ )	
	Mixture 4: NPL A431	Mixture 5: NPL 2059
<i>n</i> -hexane	983.6	-
<i>n</i> -heptane	100.7	-
<i>n</i> -octane	37.8	-
<i>n</i> -nonane	10.5	-
<i>n</i> -decane	5.8	-
Hydrogen sulphide	-	151.4
Ethyl methyl sulphide	-	103.3
Ethyl methyl sulphide	-	102.1
Dimethyl sulphide	-	102.0
Methanethiol	-	100.4
Ethanethiol	-	98.4
Carbonyl sulphide	-	98.2
2-propanethiol	-	98.1
Diethyl sulphide	-	96.1
Balance	Methane	Methane

GC-IMS chromatogram of an odorant mixture, NPL 2059 (red) overlaid with a siloxane mixture NPL3894: 20 x dilution (green).




# ISO standard - overview

The following slides give an overview of items within the proposed ISO standard

# Materials

- **Calibrations gases** traceable to a national standard, certified in accordance with ISO 6142, ISO 6143 or ISO 6144.



- **Diluent gas** (for dynamic calibration only),  $\geq 99.999\%$   $\text{CH}_4$



- **Carrier & drift gas** ( $\geq 99.999\%$   $\text{N}_2$ )





# Apparatus

- GC-IMS analyser
- Pressure regulator
- Auxiliary valves, tubing & accessories

# Sampling (refer also to ISO 10715)

- Siloxanes are highly adsorptive
- Sampling considerations should be of high importance
- Temperature should be controlled to prevent condensation of samples and calibration gases
- Appropriate sampling materials should be used
- Connections should be kept clean (free of debris and contaminants)

# Sampling into vessels

- Appropriate for use when biomethane is at or above atmospheric pressure
- Source temperature can be greater or less than sample container.
- Use of evacuated vessel (e.g. gas cylinder or sample cannister)
- Wetted surfaces should be of appropriate material or passivation to prevent adsorption

# Installing the calibration gas cylinder

- Minimise flow path surface area by installing close to the analyser
- Minimise number of connections



# Pressure control

- May require **sample pump** if sample pressure is close to atmospheric
- Ensure pressure regulators are of appropriate **material & passivation**
- Do not use a calibration mixture below the stated **certificate pressure**
- Use the **same injection pressure** for both sample and standard

# Purging of wetted flow path

- Important to ensure accurate results are achieved for siloxane analysis
- Minimum number of “Fill & empty” cycles or continuous flow required
- Additional purging required if switching from high to low siloxane content



# Flow & diffusion control

- Flow rate recommended to be between 20 ml min<sup>-1</sup> and 50 ml min<sup>-1</sup>  
(Based on 1/8 in or 1/4 in tubing)
- Sample pump can be used to achieve this
- Non-permeable membranes in pressure regulators
- Be aware of diffusion of humidity through certain polymer materials

- Perform a multipoint calibration when:
  - the analyser is first installed;
  - the analyser has had maintenance that could affect its response characteristics;
  - the analyser shows drift in excess of performance specifications as determined via comparison with a calibration standard.

## Multiple reference standards:

- Use a series of reference standards over the desired range of interest
- Should be in accordance with **ISO 10723**

## Dynamic dilution system:

- Reference standard + CH<sub>4</sub> diluent
- Should be in accordance with **ISO 6145**

# Procedure

- Follow safety precautions as recommended by manufacturer
- Perform quantitative analysis in accordance with **ISO 6143**

# Conclusion

- GC-IMS is a suitable technique for performing quantitative analysis of siloxane content within biomethane.
- It is capable of achieving measurement within the EN 16723 total Si values.



Department for  
Business, Energy  
& Industrial Strategy

**FUNDED BY BEIS**

Funding for this work has been provided EURAMETs EMPIR programme, funded by the European Union's Horizon 2020 research and innovation programme and Participating EMPIR States

The National Physical Laboratory is operated by NPL Management Ltd, a wholly-owned company of the Department for Business, Energy and Industrial Strategy (BEIS).