

# 18SIP03: Improvement of the European quality infrastructure for the measurement of total silicon and sulphur content of biogas


Project workshop 24/11/2021

**Lucy Culleton**  
*National Physical Laboratory*



# Project overview

To improve the quality infrastructure for the measurement of total silicon and sulphur content of biogas



# Project objectives

- To evaluate the measurement capability of industrial laboratories performing measurement of total silicon and total sulphur concentration in biomethane with traceable primary gas standards, and to publish the results in an industry publication or an open-access peer-reviewed journal.
- To disseminate knowledge outputs of EMRP project ENG54 Metrology for biogas in best practice sampling and analysis of total silicon and total sulphur concentration of biomethane to industrial analysis laboratories.
- To increase the awareness of standards EN 16723-1 and EN 16723-2 within the wider biogas and biomethane industry to support their wider uptake, and to provide input to ISO working group ISO/TC193/SC1/WG25 Biomethane.

# Compositions

## Siloxanes comparison

**Table 1: Nominal ranges of amount-of-substance fractions**

| Component                         | Chemical formula      | Mixture nominal amount fractions<br><i>x (mol/mol)</i> |
|-----------------------------------|-----------------------|--|
| Hexamethyldisiloxane (L2)         | $C_6H_{18}OSi_2$      | Between $(50 - 300) \cdot 10^{-9}$                     |
| Octamethylcyclotetrasiloxane (D4) | $C_8H_{24}O_4Si_4$    | Between $(50 - 300) \cdot 10^{-9}$                     |
| Decamethylcyclopentasiloxane (D5) | $C_{10}H_{30}O_5Si_5$ | Between $(50 - 300) \cdot 10^{-9}$                     |
| Methane                           | $CH_4$                | Balance  |

## Sulphur comparison

| Component                  | Chemical formula | Mixture nominal amount fractions<br><i>x (mol/mol)</i> |
|----------------------------|------------------|--|
| Hydrogen sulphide          | $H_2S$           | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Carbonyl sulphide          | $COS$            | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Methyl mercaptan           | $CH_3SH$         | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Ethyl mercaptan            | $C_2H_5SH$       | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Tetrahydrothiophene (THT)  | $C_4H_8S$        | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Dimethyl sulphide (DMS)    | $CH_3SCH_3$      | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Diethyl sulphide (DES)     | $(C_2H_5)_2S$    | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Dimethyl disulphide (DMDS) | $CH_3SSCH_3$     | Between $(1.0 - 10.0) \cdot 10^{-6}$                   |
| Methane                    | $CH_4$           | Balance  |

# Siloxanes in Biomethane




EC Mandate M/475

- 20% of energy from **renewable** sources
- 10% of transport from **biofuels**

TC 408 (biomethane)

- **EN 16723-1** (grid injection)
  - **EN 16723-2** (automotive)
- 
- National biomethane quality standardisation



# Support for Impact Project

## Total Silicon Round Robin



**Aim: To evaluate the measurement capabilities of silicon-containing compounds in methane gas standards by a round robin test.**

### Partners:




# Support for Impact Project

- The National Physical Laboratory has successfully established novel methods for **high accuracy** preparation and analysis of **traceable reference gas mixtures** containing siloxanes (L2, D4 and D5) in methane.




# Support for Impact Project

- 9 participating laboratories (including NPL)






# Standard Preparation




Gas standards prepared to a total silicon mass fraction of  $1.5 \text{ mg/m}^3$  via multi-stage dilution



# Standard Preparation



- The gas mixtures were prepared by NPL in accordance with ISO 6142-1<sup>3</sup> (gravimetric method) in **high-pressure passivated cylinders**.



- Bespoke custom made ‘**micro-loops**’ used to prepare a ‘**parent**’ mixture from pure liquid siloxanes, which is diluted to the amount-of-substance fractions.

# Standard Validation


- The comparison mixtures were assigned a value in accordance with ISO 6143<sup>4</sup> using a **direct comparison method**.




- A Gas Chromatograph system fitted with a **Flame Ionisation Detector** and a **Mass Selective Detector** was used at NPL to determine amount fractions.

# Standard Validation

## FID




## MS



# Standard re-verification

- Some siloxanes have been known to exhibit stability challenges with certain cylinder passivation in previous projects (e.g. L3/L2 siloxane).<sup>5</sup>

**Siloxane stability in cylinder passivation type A**



- The re-verification of comparison mixtures returned to NPL has indicated **stability in L2, D4 and D5 siloxanes** amount fractions within the assigned uncertainties.

# Measurement technique summary

| Participant code | Analytical Technique   | Calibration  |
|------------------|--|--|
| NPL              | <b>GC-MS</b><br><i>(Gas Chromatography Mass Spectrometry)</i>  | Gas standard prepared. Sampling by direct gas injection. Single-point calibration according to ISO 12963.  |
| L01              | <b>TD-GC-MS</b><br><i>(Thermal Desorption Gas Chromatography Mass Spectrometry)</i>                      | Liquid standard prepared in methanol. Sampling in thermal desorption tubes following ISO 16017-1.          |
| L02              | <b>GC-MS</b><br><i>(Gas Chromatography Mass Spectrometry)</i>  | Reference solution prepared. Sampling in nitrogen filled Tedlar bags. Single-point calibration.            |
| L03              | <b>GC-MS</b><br><i>(Gas Chromatography Mass Spectrometry)</i>  | Liquid standard prepared in toluene. Sampling in methane filled Tedlar bags. Three-point calibration.      |
| L04              | <b>GC-IMS</b><br><i>(Gas Chromatography Ion Mobility Spectrometry)</i>                                   | Gas standard prepared by automatic gas-dilution. Sampling in Certified permeation tubes.                   |
| L05              | <b>GC-FID</b><br><i>(Gas Chromatography Flame Ionisation Detection)</i>                                  | Liquid standard prepared. Sampling in nitrogen filled Tedlar bags.   |
| L06              | <b>ATD-GC-FID</b><br><i>(Automated Thermal Desorption Gas Chromatography Flame Ionisation Detection)</i> | Gas standard prepared according to ISO 6142-1. Sampling in thermal desorption tubes following ISO 16017-1. |
| L07              | <b>GC-ICP-MS</b><br><i>(Gas Chromatography Inductively Coupled Plasma Mass Spectrometry)</i>             | Liquid standard prepared in isopropanol. Sampling by "liquid quench system".                               |
| L08              | <b>GC-AED</b><br><i>(Gas Chromatography Atomic Emission Detection)</i>                                   | Gas standard. Sampling by direct gas injection following ASTM D8230.                                       |
| L09              | <b>GC-MS</b><br><i>(Gas Chromatography Mass Spectrometry)</i>  | Liquid standard prepared. Sampling in sorbent tube following ASTM D8230.                                   |

# Comparison criteria

$$Z_i = \frac{(x_i - X_{ref})}{\hat{\sigma}}$$

- $Z_i$  denotes the z score of laboratory  $i$
- $x_i$  denotes the reported result of laboratory  $i$
- $X_{ref}$  denotes the reference value of the mixture assigned to lab  $i$
- $\hat{\sigma}$  denotes the standard deviation for proficiency assessment, defined for this comparison as double the uncertainty in the reference values, equating to 16% of  $X_{ref}$  (0.035 for L2, 0.018 for D4, 0.013 for D5 and 0.248 for total silicon for labs NPL, L01, L02, L03, L04 and L06 and 0.250 for labs L05, L07, L08 and L09)

Interpretation:

- $|z| \leq 2$  Satisfactory result (shown in green within tables)
- $2 < |z| < 3$  Questionable result (shown in orange within tables)
- $|z| > 3$  Unsatisfactory result (shown in red within tables)

# Comparison criteria

$$E_{n,i} = \frac{x_i - X_{ref}}{\sqrt{(U_i)^2 + (U_{ref})^2}}$$

- $E_{n,i}$  denotes the  $E_n$ -score of laboratory  $i$
- $x_i$  denotes the reported result of laboratory  $i$
- $x_{ref}$  denotes the reference value
- $U_i$  denotes the expanded uncertainty (with coverage factor  $k=2$ ) given by laboratory  $i$
- $U_{ref}$  denotes the expanded uncertainty (with coverage factor  $k=2$ ) of the reference value

Interpretation:

$$|E_n| \leq 1$$

Satisfactory result (shown in green within tables)


$$|E_n| > 1$$

Unsatisfactory result (shown in red within tables)



# D5 results

## Reported amount fractions for D5 Siloxane




| Participant code | Analytical Technique | Z-score | Quality of result | $E_n$ number | Quality of result |
|------------------|----------------------|---------|-------------------|--------------|-------------------|
| NPL              | GC-MS                | 0.20    | Satisfactory      | 0.39         | Satisfactory      |
| L01              | TD-GC-MSD            | 0.90    | Satisfactory      | 0.54         | Satisfactory      |
| L02              | GC-MS                | -0.23   | Satisfactory      | -0.29        | Satisfactory      |
| L03              | GC-MS                | 3.51    | Unsatisfactory    | 1.11         | Unsatisfactory    |
| L04              | GC-IMS               | -0.14   | Satisfactory      | -0.27        | Satisfactory      |
| L05              | GC-FID               | 1.23    | Satisfactory      | 0.99         | Satisfactory      |
| L06              | ATD-GC-FID           | 0.06    | Satisfactory      | 0.08         | Satisfactory      |
| L07              | GC-ICP-MS            | 0.06    | Satisfactory      | 0.11         | Satisfactory      |
| L08              | GC-AED               | -0.23   | Satisfactory      | -0.25        | Satisfactory      |
| L09              | GC-MS                | -1.73   | Satisfactory      | -2.97        | Unsatisfactory    |

### Key points:

- L09 noted issues (incorrect sampling volume used for sorbent tubes resulting in low recovery)

# D4 results

## Reported amount fractions for D4 Siloxane




| Participant code | Analytical Technique | Z-score | Quality of result | $E_n$ number | Quality of result |
|------------------|----------------------|---------|-------------------|--------------|-------------------|
| NPL              | GC-MS                | 0.00    | Satisfactory      | -0.01        | Satisfactory      |
| L01              | TD-GC-MSD            | 1.49    | Satisfactory      | 1.14         | Unsatisfactory    |
| L02              | GC-MS                | -0.11   | Satisfactory      | -0.14        | Satisfactory      |
| L03              | GC-MS                | -0.27   | Satisfactory      | -0.15        | Satisfactory      |
| L04              | GC-IMS               | 0.61    | Satisfactory      | 1.19         | Unsatisfactory    |
| L05              | GC-FID               | -0.02   | Satisfactory      | -0.02        | Satisfactory      |
| L06              | ATD-GC-FID           | 0.06    | Satisfactory      | 0.08         | Satisfactory      |
| L07              | GC-ICP-MS            | -0.01   | Satisfactory      | -0.02        | Satisfactory      |
| L08              | GC-AED               | 0.94    | Satisfactory      | 0.77         | Satisfactory      |
| L09              | GC-MS                | -1.27   | Satisfactory      | -1.71        | Unsatisfactory    |

### Key points:

- Best agreement
- L09 noted issues (incorrect sampling volume used for sorbent tubes resulting in low recovery)

# L2 results

## Reported amount fractions for L2 Siloxane




| Participant code | Analytical Technique | Z-score | Quality of result | $E_n$ number | Quality of result |
|------------------|----------------------|---------|-------------------|--------------|-------------------|
| NPL              | GC-MS                | 0.06    | Satisfactory      | 0.12         | Satisfactory      |
| L01              | TD-GC-MSD            | 2.29    | Questionable      | 1.68         | Unsatisfactory    |
| L02              | GC-MS                | 1.22    | Satisfactory      | 1.35         | Unsatisfactory    |
| L03              | GC-MS                | -1.42   | Satisfactory      | -1.62        | Unsatisfactory    |
| L04              | GC-IMS               | -2.29   | Questionable      | -3.98        | Unsatisfactory    |
| L05              | GC-FID               | -1.67   | Satisfactory      | -2.84        | Unsatisfactory    |
| L06              | ATD-GC-FID           | 0.17    | Satisfactory      | 0.27         | Satisfactory      |
| L07              | GC-ICP-MS            | -0.40   | Satisfactory      | -0.73        | Satisfactory      |
| L08              | GC-AED               | -1.19   | Satisfactory      | -1.57        | Unsatisfactory    |
| L09              | GC-MS                | -2.88   | Questionable      | -5.00        | Unsatisfactory    |

### Key points:

- Poorest agreement of all siloxanes
- L04 noted the permeation tube used had passed its expiry date
- L08 notes issues with GC-AED (solved after comparison-carbon in reagent O<sub>2</sub>)
- L09 noted issues (incorrect sampling volume used for sorbent tubes resulting in low recovery)

# Total silicon results

## Reported mass fractions for Total Silicon



| Participant code | Analytical Technique | Z-score | Quality of result | E <sub>n</sub> number | Quality of result |
|------------------|----------------------|---------|-------------------|-----------------------|-------------------|
| NPL              | GC-MS                | -       | -                 | -                     | -                 |
| L01              | TD-GC-MSD            | 1.61    | Satisfactory      | 0.97                  | Satisfactory      |
| L02              | GC-MS                | 0.69    | Satisfactory      | 0.80                  | Satisfactory      |
| L03              | GC-MS                | -1.05   | Satisfactory      | -0.43                 | Satisfactory      |
| L04              | GC-IMS               | -0.85   | Satisfactory      | -                     | -                 |
| L05              | GC-FID               | 0.16    | Satisfactory      | 0.25                  | Satisfactory      |
| L06              | ATD-GC-FID           | -       | -                 | -                     | -                 |
| L07              | GC-ICP-MS            | -       | -                 | -                     | -                 |
| L08              | GC-AED               | -0.24   | Satisfactory      | -0.18                 | Satisfactory      |
| L09              | GC-MS                | -1.44   | Satisfactory      | -1.53                 | Unsatisfactory    |

### Key points:

- Total silicon not reported by all labs
- Generally good agreement

# Summary

- Final report to be **distributed this week**
  - Generally good agreement for siloxanes and total silicon, with notable issues for L2 siloxane
  - Further study recommended to ascertain whether differences are due to **calibration, sampling or the measurement method**
- e.g.
- Use of the same sampling and analysis method, testing **different calibration techniques**
  - Use of the same calibration and analysis method, testing **different sampling techniques.**
  - Use of the same calibration and sampling techniques, testing **different analysis methods.**
- Expand on number of siloxanes and investigate influences of likely interferants